

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016838.D
 Acq On : 23 Dec 2023 04:44
 Operator : SY/MD
 Sample : 05957-10
 Misc : 3.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM46

Quant Time: Dec 23 05:36:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.826	168	104772	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	233558	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	232152	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	87698	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	64741	67.688	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 135.380%	
35) Dibromofluoromethane	7.752	113	70468	52.671	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.340%	
50) Toluene-d8	10.209	98	258735	53.647	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.300%	
62) 4-Bromofluorobenzene	12.508	95	82527	52.743	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	= 105.480%	
Target Compounds						
20) Methylene Chloride	4.765	84	10361	5.076	ug/l	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016838.D
Acq On : 23 Dec 2023 04:44
Operator : SY/MD
Sample : 05957-10
Misc : 3.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM46

Quant Time: Dec 23 05:36:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.826 min Scan# 91
Delta R.T. 0.025 min
Lab File: VY016838.D
Acq: 23 Dec 2023 04:44




Instrument :
MSVOA_Y
ClientSampleId :
COM46


Tgt Ion:168 Resp: 104772
Ion Ratio Lower Upper
168 100
99 54.4 44.2 66.4



#20
Methylene Chloride
Concen: 5.076 ug/l
RT: 4.765 min Scan# 483
Delta R.T. 0.031 min
Lab File: VY016838.D
Acq: 23 Dec 2023 04:44

Tgt Ion: 84 Resp: 10361
Ion Ratio Lower Upper
84 100
49 118.3 95.2 142.8
51 34.1 29.9 44.9
86 64.8 52.2 78.4





#33

1,2-Dichloroethane-d4

Concen: 67.688 ug/l

RT: 8.173 min Scan# 110

Delta R.T. 0.019 min

Lab File: VY016838.D

Acq: 23 Dec 2023 04:44

Instrument :

MSVOA_Y

ClientSampleId :





COM46

Tgt Ion: 65 Resp: 64741

Ion Ratio Lower Upper

65 100

67 53.7 0.0 109.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.722 min Scan# 1132

Delta R.T. 0.019 min

Lab File: VY016838.D

Acq: 23 Dec 2023 04:44




Tgt Ion: 114 Resp: 233558


Ion Ratio Lower Upper

114 100

63 20.8 0.0 37.4

88 14.8 0.0 29.0





#35

Dibromofluoromethane

Concen: 52.671 ug/l

RT: 7.752 min Scan# 91

Delta R.T. 0.025 min

Lab File: VY016838.D

Acq: 23 Dec 2023 04:44

Instrument :

MSVOA_Y

ClientSampleId :

COM46





Tgt Ion:113 Resp: 70468

Ion Ratio Lower Upper

113 100

111 103.6 83.8 125.6

192 17.6 14.5 21.7



#50

Toluene-d8

Concen: 53.647 ug/l

RT: 10.209 min Scan# 1376

Delta R.T. 0.019 min

Lab File: VY016838.D




Acq: 23 Dec 2023 04:44


Tgt Ion: 98 Resp: 258735

Ion Ratio Lower Upper

98 100

100 65.4 52.3 78.5









#62
4-Bromofluorobenzene
Concen: 52.743 ug/l
RT: 12.508 min Scan# 11
Delta R.T. 0.019 min
Lab File: VY016838.D
Acq: 23 Dec 2023 04:44




Instrument :
MSVOA_Y
ClientSampleId :
COM46


Tgt Ion: 95 Resp: 82527
Ion Ratio Lower Upper
95 100
174 77.2 0.0 160.0
176 75.4 0.0 151.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.514 min Scan# 1590
Delta R.T. 0.012 min
Lab File: VY016838.D
Acq: 23 Dec 2023 04:44

Tgt Ion: 117 Resp: 232152
Ion Ratio Lower Upper
117 100
82 55.7 43.8 65.8
119 31.3 26.5 39.7





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.447 min Scan# 1907

Delta R.T. 0.012 min

Lab File: VY016838.D


Acq: 23 Dec 2023 04:44

Instrument :

MSVOA_Y

ClientSampleId :

COM46



Tgt Ion:152 Resp: 87698

Ion Ratio Lower Upper

152 100

115 59.7 30.0 90.0

150 155.5 0.0 344.6

