

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122322\
 Data File : VY011991.D
 Acq On : 23 Dec 2022 12:41
 Operator : KP/MD
 Sample : N6196-09
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 23 22:24:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	192176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	335204	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	310320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	140208	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	119545	61.744	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.480%
35) Dibromofluoromethane	7.716	113	104620	51.843	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.680%
50) Toluene-d8	10.179	98	400756	48.947	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.900%
62) 4-Bromofluorobenzene	12.483	95	127924	46.456	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122322\
Data File : VY011991.D
Acq On : 23 Dec 2022 12:41
Operator : KP/MD
Sample : N6196-09
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 23 22:24:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

