

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122322\
 Data File : VY012006.D
 Acq On : 23 Dec 2022 18:32
 Operator : KP/MD
 Sample : N6098-11
 Misc : 6.19g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FVSUB-TP2-COMP-02

Quant Time: Dec 23 22:27:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	192501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	339658	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	318929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	143186	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	115969	59.796	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.600%
35) Dibromofluoromethane	7.722	113	106584	52.123	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.240%
50) Toluene-d8	10.179	98	412326	49.699	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.483	95	129940	46.569	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122322\
Data File : VY012006.D
Acq On : 23 Dec 2022 18:32
Operator : KP/MD
Sample : N6098-11
Misc : 6.19g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FVSUB-TP2-COMP-02

Quant Time: Dec 23 22:27:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

