



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122322\
 Data File : VY011986.D
 Acq On : 23 Dec 2022 10:05
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Title : SW846 8260
 Last Update : Tue Dec 20 12:37:37 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	12399	PASS
75	95	30	60	51.6	36312	PASS
95	95	100	100	100.0	70336	PASS
96	95	5	9	6.7	4739	PASS
173	174	0.00	2	1.1	687	PASS
174	95	50	100	88.6	62344	PASS
175	174	5	9	6.9	4332	PASS
176	174	95	101	95.4	59485	PASS
177	176	5	9	7.1	4243	PASS