

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020690.D
 Acq On : 23 Dec 2024 13:46
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 23 14:31:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 23 14:10:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	122966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	172651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	141587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	67269	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	53847	49.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.880%
35) Dibromofluoromethane	7.646	113	49110	48.443	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.880%
50) Toluene-d8	10.115	98	193162	52.460	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.920%
62) 4-Bromofluorobenzene	12.414	95	61955	46.229	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	92.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020690.D
Acq On : 23 Dec 2024 13:46
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 23 14:31:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 23 14:10:31 2024
Response via : Initial Calibration

