

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020697.D
 Acq On : 23 Dec 2024 17:27
 Operator : SY/MD
 Sample : P5339-01RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-12182024RE

Quant Time: Dec 25 01:13:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 25 00:58:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	243	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	378	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.426	117	143	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	179	83.174	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	166.340%#
35) Dibromofluoromethane	7.652	113	68	30.637	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.280%
50) Toluene-d8	10.115	98	321	39.819	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.640%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020697.D
Acq On : 23 Dec 2024 17:27
Operator : SY/MD
Sample : P5339-01RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-12182024RE

Quant Time: Dec 25 01:13:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 25 00:58:45 2024
Response via : Initial Calibration

