

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020700.D
 Acq On : 23 Dec 2024 18:38
 Operator : SY/MD
 Sample : P5306-11
 Misc : 5.95g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-VSL-12-121224

Quant Time: Dec 25 01:02:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 25 00:58:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	94664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	149835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	128778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	55306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	55443	66.131	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.260%
35) Dibromofluoromethane	7.646	113	48256	54.849	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.700%
50) Toluene-d8	10.115	98	185787	58.140	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.280%
62) 4-Bromofluorobenzene	12.414	95	58708	50.477	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	100.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020700.D
Acq On : 23 Dec 2024 18:38
Operator : SY/MD
Sample : P5306-11
Misc : 5.95g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-VSL-12-121224

Quant Time: Dec 25 01:02:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 25 00:58:45 2024
Response via : Initial Calibration

