

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
 Data File : VY020701.D
 Acq On : 23 Dec 2024 19:01
 Operator : SY/MD
 Sample : P5306-15
 Misc : 4.90g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-VSL-14-121224

Quant Time: Dec 25 01:02:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 25 00:58:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	94774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	152816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	128401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	54196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	54157	64.522	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.040%
35) Dibromofluoromethane	7.640	113	48833	54.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.840%
50) Toluene-d8	10.115	98	186989	57.375	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.760%
62) 4-Bromofluorobenzene	12.414	95	58946	49.693	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	99.380%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122324\
Data File : VY020701.D
Acq On : 23 Dec 2024 19:01
Operator : SY/MD
Sample : P5306-15
Misc : 4.90g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-VSL-14-121224

Quant Time: Dec 25 01:02:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122324S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 25 00:58:45 2024
Response via : Initial Calibration

