

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
 Data File : VY020714.D
 Acq On : 26 Dec 2024 13:56
 Operator : SY/MD
 Sample : VY1226SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1226SBL01

Quant Time: Dec 26 16:15:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	99549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	157271	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	133794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	56315	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	52717	64.652	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.300%
35) Dibromofluoromethane	7.640	113	49885	57.586	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.180%
50) Toluene-d8	10.109	98	181408	52.569	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.407	95	58692	50.981	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	101.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
Data File : VY020714.D
Acq On : 26 Dec 2024 13:56
Operator : SY/MD
Sample : VY1226SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1226SBL01

Quant Time: Dec 26 16:15:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

