

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
 Data File : VY020717.D
 Acq On : 26 Dec 2024 15:18
 Operator : SY/MD
 Sample : P5383-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-12232024

Quant Time: Dec 26 16:30:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	721	50.000	ug/l	#-0.01
72) 1,4-Dichlorobenzene-d4	13.359	152	100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	624	107.299	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 214.600%#	
35) Dibromofluoromethane	7.634	113	440	72.885	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 145.760%	
50) Toluene-d8	10.115	98	1322	54.893	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.780%	
62) 4-Bromofluorobenzene	12.414	95	117	14.583	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 29.160%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
Data File : VY020717.D
Acq On : 26 Dec 2024 15:18
Operator : SY/MD
Sample : P5383-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-12232024

Quant Time: Dec 26 16:30:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

