

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
 Data File : VY020719.D
 Acq On : 26 Dec 2024 16:04
 Operator : SY/MD
 Sample : P5383-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-12232024

Quant Time: Dec 26 16:30:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	3184	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.622	114	4771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	3476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	1585	60.775	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.540%
35) Dibromofluoromethane	7.640	113	1605	61.075	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.140%
50) Toluene-d8	10.109	98	5115	48.982	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.960%
62) 4-Bromofluorobenzene	12.414	95	1014	29.034	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	58.060%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122624\
Data File : VY020719.D
Acq On : 26 Dec 2024 16:04
Operator : SY/MD
Sample : P5383-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-12232024

Quant Time: Dec 26 16:30:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

