

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv122719\
 Data File : VY001088.D
 Acq On : 27 Dec 2019 11:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 28 04:15:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 23 12:07:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.436	0.422	3.2	93	0.00
3 P	Chloromethane	0.593	0.511	13.8	84	0.00
4 C	Vinyl Chloride	0.576	0.534	7.3#	90	0.00
5 T	Bromomethane	0.354	0.342	3.4	91	0.00
6 T	Chloroethane	0.359	0.350	2.5	93	0.00
7 T	Trichlorofluoromethane	0.845	0.879	-4.0	103	0.00
8 T	Diethyl Ether	0.345	0.370	-7.2	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.620	-11.7	108	0.00
10 T	Methyl Iodide	0.767	0.790	-3.0	97	0.00
11 T	Tert butyl alcohol	0.054	0.056	-3.7	103	0.00
12 CM	1,1-Dichloroethene	0.562	0.596	-6.0#	103	0.00
13 T	Acrolein	0.067	0.069	-3.0	89	0.00
14 T	Allyl chloride	0.873	0.893	-2.3	99	0.00
15 T	Acrylonitrile	0.172	0.196	-14.0	108	0.00
16 T	Acetone	0.114	0.102	10.5	93	0.00
17 T	Carbon Disulfide	1.836	1.565	14.8	81	0.00
18 T	Methyl Acetate	0.436	0.453	-3.9	105	0.00
19 T	Methyl tert-butyl Ether	1.450	1.616	-11.4	106	0.00
20 T	Methylene Chloride	0.669	0.690	-3.1	106	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.663	-6.4	102	0.00
22 T	Diisopropyl ether	1.764	1.827	-3.6	99	0.00
23 T	Vinyl Acetate	1.114	1.158	-3.9	98	0.00
24 P	1,1-Dichloroethane	1.042	1.117	-7.2	103	0.00
25 T	2-Butanone	0.200	0.200	0.0	98	0.00
26 T	2,2-Dichloropropane	0.856	0.944	-10.3	107	0.00
27 T	cis-1,2-Dichloroethene	0.684	0.771	-12.7	108	0.00
28 T	Bromochloromethane	0.407	0.389	4.4	85	0.00
29 T	Tetrahydrofuran	0.142	0.144	-1.4	100	0.00
30 C	Chloroform	0.986	1.096	-11.2#	107	0.00
31 T	Cyclohexane	1.096	1.005	8.3	93	0.00
32 T	1,1,1-Trichloroethane	0.841	0.936	-11.3	108	0.00
33 S	1,2-Dichloroethane-d4	0.493	0.473	4.1	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.294	0.285	3.1	92	0.00
36 T	1,1-Dichloropropene	0.477	0.511	-7.1	104	0.00
37 T	Ethyl Acetate	0.264	0.279	-5.7	103	0.00
38 T	Carbon Tetrachloride	0.418	0.472	-12.9	108	0.00
39 T	Methylcyclohexane	0.633	0.665	-5.1	100	0.00
40 TM	Benzene	1.439	1.587	-10.3	106	0.00
41 T	Methacrylonitrile	0.115	0.110	4.3	83	0.00
42 TM	1,2-Dichloroethane	0.344	0.371	-7.8	101	0.00
43 T	Isopropyl Acetate	0.489	0.519	-6.1	101	0.00
44 TM	Trichloroethene	0.379	0.429	-13.2	109	0.00
45 C	1,2-Dichloropropane	0.361	0.395	-9.4#	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv122719\
 Data File : VY001088.D
 Acq On : 27 Dec 2019 11:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 28 04:15:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 23 12:07:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.186	0.209	-12.4	105	0.00
47 T	Bromodichloromethane	0.431	0.489	-13.5	106	0.00
48 T	Methyl methacrylate	0.209	0.217	-3.8	97	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	110	0.00
50 S	Toluene-d8	1.151	1.201	-4.3	93	0.00
51 T	4-Methyl-2-Pentanone	0.255	0.271	-6.3	101	0.00
52 CM	Toluene	0.886	0.992	-12.0#	106	0.00
53 T	t-1,3-Dichloropropene	0.468	0.532	-13.7	106	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.629	-11.7	106	0.00
55 T	1,1,2-Trichloroethane	0.278	0.320	-15.1	110	0.00
56 T	Ethyl methacrylate	0.395	0.448	-13.4	105	0.00
57 T	1,3-Dichloropropane	0.487	0.541	-11.1	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.145	5.2	101	0.00
59 T	2-Hexanone	0.176	0.189	-7.4	102	0.00
60 T	Dibromochloromethane	0.297	0.353	-18.9	111	0.00
61 T	1,2-Dibromoethane	0.266	0.308	-15.8	108	0.00
62 S	4-Bromofluorobenzene	0.421	0.384	8.8	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.433	0.491	-13.4	108	0.00
65 PM	Chlorobenzene	1.037	1.184	-14.2	109	0.00
66 T	1,1,1,2-Tetrachloroethane	0.342	0.403	-17.8	108	0.00
67 C	Ethyl Benzene	1.865	2.142	-14.9#	107	0.00
68 T	m/p-Xylenes	0.708	0.817	-15.4	108	0.00
69 T	o-Xylene	0.666	0.764	-14.7	107	0.00
70 T	Styrene	1.133	1.329	-17.3	108	0.00
71 P	Bromoform	0.191	0.227	-18.8	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.869	4.548	-17.5	107	0.00
74 T	N-amyl acetate	1.016	1.084	-6.7	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.902	-18.4	107	0.00
76 T	1,2,3-Trichloropropane	0.616	0.628	-1.9	91	0.00
77 T	Bromobenzene	0.870	1.024	-17.7	108	0.00
78 T	n-propylbenzene	4.715	5.530	-17.3	108	0.00
79 T	2-Chlorotoluene	2.604	2.977	-14.3	105	0.00
80 T	1,3,5-Trimethylbenzene	3.172	3.679	-16.0	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.289	0.352	-21.8	107	0.00
82 T	4-Chlorotoluene	2.704	3.036	-12.3	102	0.00
83 T	tert-Butylbenzene	2.662	3.219	-20.9	111	0.00
84 T	1,2,4-Trimethylbenzene	3.187	3.700	-16.1	105	0.00
85 T	sec-Butylbenzene	3.925	4.590	-16.9	107	0.00
86 T	p-Isopropyltoluene	3.396	4.024	-18.5	107	0.00
87 T	1,3-Dichlorobenzene	1.671	1.942	-16.2	106	0.00
88 T	1,4-Dichlorobenzene	1.686	1.949	-15.6	106	0.00
89 T	n-Butylbenzene	3.417	3.969	-16.2	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv122719\
Data File : VY001088.D
Acq On : 27 Dec 2019 11:09
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 28 04:15:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 23 12:07:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.640	0.771	-20.5	106	0.00
91 T	1,2-Dichlorobenzene	1.542	1.782	-15.6	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.138	-7.8	98	0.00
93 T	1,2,4-Trichlorobenzene	1.041	1.195	-14.8	104	0.00
94 T	Hexachlorobutadiene	0.495	0.578	-16.8	106	0.00
95 T	Naphthalene	2.447	2.797	-14.3	103	0.00
96 T	1,2,3-Trichlorobenzene	0.928	1.044	-12.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6