

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122722\
 Data File : VY012015.D
 Acq On : 27 Dec 2022 14:12
 Operator : KP/MD
 Sample : N6190-01RE
 Misc : 6.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LOCATION-1RE

Quant Time: Dec 28 00:38:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	2301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	2720	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	1628	158.423	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	316.840%#
35) Dibromofluoromethane	7.710	113	474	34.217	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	68.440%
50) Toluene-d8	10.185	98	3007	53.502	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.000%
62) 4-Bromofluorobenzene	12.483	95	1113	58.881	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122722\
Data File : VY012015.D
Acq On : 27 Dec 2022 14:12
Operator : KP/MD
Sample : N6190-01RE
Misc : 6.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LOCATION-1RE

Quant Time: Dec 28 00:38:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

