

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122722\
 Data File : VY012012.D
 Acq On : 27 Dec 2022 13:02
 Operator : KP/MD
 Sample : N6135-02
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11977

Quant Time: Dec 27 13:38:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	181518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	317096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	293373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	127067	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	105906	57.912	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.820%
35) Dibromofluoromethane	7.722	113	97443	51.044	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.080%
50) Toluene-d8	10.179	98	382401	49.372	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.483	95	118292	45.411	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122722\
Data File : VY012012.D
Acq On : 27 Dec 2022 13:02
Operator : KP/MD
Sample : N6135-02
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11977

Quant Time: Dec 27 13:38:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY012012.D
Acq: 27 Dec 2022 13:02




Instrument :
MSVOA_Y
ClientSampleId :
72-11977


Tgt Ion:168 Resp: 181518
Ion Ratio Lower Upper
168 100
99 55.5 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 57.912 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY012012.D
Acq: 27 Dec 2022 13:02

Tgt Ion: 65 Resp: 105906
Ion Ratio Lower Upper
65 100
67 51.1 0.0 103.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012012.D

Acq: 27 Dec 2022 13:02

Instrument :

MSVOA_Y

ClientSampleId :

72-11977





Tgt Ion:114 Resp: 317096

Ion Ratio Lower Upper

114 100

63 18.9 0.0 39.6

88 15.3 0.0 32.2



#35

Dibromofluoromethane

Concen: 51.044 ug/l

RT: 7.722 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY012012.D

Acq: 27 Dec 2022 13:02




Tgt Ion:113 Resp: 97443


Ion Ratio Lower Upper

113 100

111 102.7 83.0 124.4

192 20.7 16.0 24.0









#50
Toluene-d8
Concen: 49.372 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY012012.D
Acq: 27 Dec 2022 13:02




Instrument :
MSVOA_Y
ClientSampleId :
72-11977


Tgt Ion: 98 Resp: 382401
Ion Ratio Lower Upper
98 100
100 64.0 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 45.411 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012012.D
Acq: 27 Dec 2022 13:02

Tgt Ion: 95 Resp: 118292
Ion Ratio Lower Upper
95 100
174 89.7 0.0 172.0
176 87.4 0.0 167.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY012012.D

Acq: 27 Dec 2022 13:02

Instrument :

MSVOA_Y

ClientSampleId :

72-11977





Tgt Ion:117 Resp: 293373

Ion Ratio Lower Upper

117 100

82 54.0 44.8 67.2

119 32.2 25.4 38.2



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VY012012.D

Acq: 27 Dec 2022 13:02

Tgt Ion:152 Resp: 127067

Ion Ratio Lower Upper

152 100

115 57.0 29.1 87.3

150 155.6 0.0 342.6

