

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
 Data File : VY020735.D
 Acq On : 27 Dec 2024 13:18
 Operator : SY/MD
 Sample : P5382-03
 Misc : 5.69g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Dec 28 03:02:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	107608	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.616	114	174360	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	145577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	55916	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	54118	61.400	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.800%
35) Dibromofluoromethane	7.640	113	55015	57.283	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.560%
50) Toluene-d8	10.109	98	201262	52.605	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.220%
62) 4-Bromofluorobenzene	12.408	95	59987	46.999	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
Data File : VY020735.D
Acq On : 27 Dec 2024 13:18
Operator : SY/MD
Sample : P5382-03
Misc : 5.69g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Dec 28 03:02:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

