

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
 Data File : VY020739.D
 Acq On : 27 Dec 2024 15:10
 Operator : SY/MD
 Sample : P5387-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-122624

Quant Time: Dec 28 03:03:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.738	168	71	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.616	114	214	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	120	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	125	214.941	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 429.880%#	
35) Dibromofluoromethane	7.640	113	67	56.840	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.680%	
50) Toluene-d8	10.109	98	119	26.236	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 52.480%#	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	= 0.000%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
Data File : VY020739.D
Acq On : 27 Dec 2024 15:10
Operator : SY/MD
Sample : P5387-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-122624

Quant Time: Dec 28 03:03:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

