

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
 Data File : VY020744.D
 Acq On : 27 Dec 2024 17:07
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 28 03:05:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	117560	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.616	114	194142	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	169786	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	70664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	61754	64.132	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.260%	
35) Dibromofluoromethane	7.640	113	59900	56.015	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.020%	
50) Toluene-d8	10.109	98	224164	52.621	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 105.240%	
62) 4-Bromofluorobenzene	12.408	95	71974	50.645	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 101.300%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.042	105	4397	1.108	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
Data File : VY020744.D
Acq On : 27 Dec 2024 17:07
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 28 03:05:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

