

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
 Data File : VY020748.D
 Acq On : 27 Dec 2024 18:41
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 28 03:06:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	118941	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.616	114	194423	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	169474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	73075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	63007	64.673	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.340%
35) Dibromofluoromethane	7.640	113	60739	56.717	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.440%
50) Toluene-d8	10.109	98	224782	52.687	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.380%
62) 4-Bromofluorobenzene	12.408	95	72114	50.670	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	101.340%
Target Compounds						
30) Chloroform	7.421	83	2091	1.121	ug/l #	84
51) 4-Methyl-2-Pentanone	10.006	43	965	2.018	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122724\
Data File : VY020748.D
Acq On : 27 Dec 2024 18:41
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 28 03:06:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

