

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv123019\
 Data File : VY001117.D
 Acq On : 30 Dec 2019 10:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 31 05:39:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 23 12:07:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.436	0.384	11.9	86	0.00
3 P	Chloromethane	0.593	0.385	35.1#	65	0.00
4 C	Vinyl Chloride	0.576	0.403	30.0#	69	0.00
5 T	Bromomethane	0.354	0.272	23.2	74	0.00
6 T	Chloroethane	0.359	0.286	20.3	77	0.00
7 T	Trichlorofluoromethane	0.845	0.820	3.0	97	0.00
8 T	Diethyl Ether	0.345	0.342	0.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.563	-1.4	100	0.00
10 T	Methyl Iodide	0.767	0.760	0.9	95	0.00
11 T	Tert butyl alcohol	0.054	0.048	11.1	89	0.00
12 CM	1,1-Dichloroethene	0.562	0.554	1.4#	97	0.00
13 T	Acrolein	0.067	0.063	6.0	83	0.00
14 T	Allyl chloride	0.873	0.756	13.4	85	0.00
15 T	Acrylonitrile	0.172	0.163	5.2	91	0.00
16 T	Acetone	0.114	0.091	20.2	84	0.00
17 T	Carbon Disulfide	1.836	1.435	21.8	76	0.00
18 T	Methyl Acetate	0.436	0.373	14.4	88	0.00
19 T	Methyl tert-butyl Ether	1.450	1.447	0.2	96	0.00
20 T	Methylene Chloride	0.669	0.654	2.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.622	0.2	97	0.01
22 T	Diisopropyl ether	1.764	1.594	9.6	88	0.00
23 T	Vinyl Acetate	1.114	0.978	12.2	84	0.00
24 P	1,1-Dichloroethane	1.042	1.013	2.8	95	0.00
25 T	2-Butanone	0.200	0.164	18.0	81	0.00
26 T	2,2-Dichloropropane	0.856	0.849	0.8	98	0.00
27 T	cis-1,2-Dichloroethene	0.684	0.717	-4.8	102	0.00
28 T	Bromochloromethane	0.407	0.360	11.5	79	0.00
29 T	Tetrahydrofuran	0.142	0.114	19.7	80	0.00
30 C	Chloroform	0.986	1.016	-3.0#	101	0.00
31 T	Cyclohexane	1.096	0.876	20.1	83	0.00
32 T	1,1,1-Trichloroethane	0.841	0.873	-3.8	103	0.00
33 S	1,2-Dichloroethane-d4	0.493	0.421	14.6	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.294	0.264	10.2	86	0.00
36 T	1,1-Dichloropropene	0.477	0.466	2.3	95	0.00
37 T	Ethyl Acetate	0.264	0.226	14.4	84	0.00
38 T	Carbon Tetrachloride	0.418	0.439	-5.0	101	0.00
39 T	Methylcyclohexane	0.633	0.620	2.1	94	0.00
40 TM	Benzene	1.439	1.475	-2.5	99	0.00
41 T	Methacrylonitrile	0.115	0.110	4.3	85	0.00
42 TM	1,2-Dichloroethane	0.344	0.346	-0.6	95	0.00
43 T	Isopropyl Acetate	0.489	0.428	12.5	84	0.00
44 TM	Trichloroethene	0.379	0.402	-6.1	103	0.00
45 C	1,2-Dichloropropane	0.361	0.358	0.8#	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv123019\
 Data File : VY001117.D
 Acq On : 30 Dec 2019 10:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 31 05:39:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 23 12:07:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.186	0.192	-3.2	97	0.00
47 T	Bromodichloromethane	0.431	0.450	-4.4	98	0.00
48 T	Methyl methacrylate	0.209	0.185	11.5	83	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	99	0.00
50 S	Toluene-d8	1.151	1.073	6.8	84	0.00
51 T	4-Methyl-2-Pentanone	0.255	0.222	12.9	84	0.00
52 CM	Toluene	0.886	0.911	-2.8#	98	0.00
53 T	t-1,3-Dichloropropene	0.468	0.485	-3.6	97	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.580	-3.0	99	0.00
55 T	1,1,2-Trichloroethane	0.278	0.289	-4.0	100	0.00
56 T	Ethyl methacrylate	0.395	0.392	0.8	92	0.00
57 T	1,3-Dichloropropane	0.487	0.493	-1.2	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.122	20.3	86	0.00
59 T	2-Hexanone	0.176	0.154	12.5	84	0.00
60 T	Dibromochloromethane	0.297	0.326	-9.8	103	0.00
61 T	1,2-Dibromoethane	0.266	0.274	-3.0	97	0.00
62 S	4-Bromofluorobenzene	0.421	0.339	19.5	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.433	0.457	-5.5	102	0.00
65 PM	Chlorobenzene	1.037	1.111	-7.1	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.342	0.383	-12.0	104	0.00
67 C	Ethyl Benzene	1.865	1.954	-4.8#	99	0.00
68 T	m/p-Xylenes	0.708	0.765	-8.1	103	0.00
69 T	o-Xylene	0.666	0.724	-8.7	103	0.00
70 T	Styrene	1.133	1.245	-9.9	102	0.00
71 P	Bromoform	0.191	0.216	-13.1	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.869	4.251	-9.9	100	0.00
74 T	N-amyl acetate	1.016	0.931	8.4	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.789	-3.5	94	0.00
76 T	1,2,3-Trichloropropane	0.616	0.581	5.7	84	0.00
77 T	Bromobenzene	0.870	0.978	-12.4	104	0.00
78 T	n-propylbenzene	4.715	5.160	-9.4	101	0.00
79 T	2-Chlorotoluene	2.604	2.779	-6.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.172	3.468	-9.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.289	0.307	-6.2	94	0.00
82 T	4-Chlorotoluene	2.704	2.870	-6.1	97	0.00
83 T	tert-Butylbenzene	2.662	2.990	-12.3	104	0.00
84 T	1,2,4-Trimethylbenzene	3.187	3.445	-8.1	99	0.00
85 T	sec-Butylbenzene	3.925	4.324	-10.2	101	0.00
86 T	p-Isopropyltoluene	3.396	3.803	-12.0	102	0.00
87 T	1,3-Dichlorobenzene	1.671	1.850	-10.7	101	0.00
88 T	1,4-Dichlorobenzene	1.686	1.871	-11.0	102	0.00
89 T	n-Butylbenzene	3.417	3.708	-8.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv123019\
Data File : VY001117.D
Acq On : 30 Dec 2019 10:47
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 31 05:39:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y122319S.M
Quant Title : SW846 8260
QLast Update : Mon Dec 23 12:07:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.640	0.727	-13.6	101	0.00
91 T	1,2-Dichlorobenzene	1.542	1.702	-10.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.124	3.1	88	0.00
93 T	1,2,4-Trichlorobenzene	1.041	1.155	-11.0	101	0.00
94 T	Hexachlorobutadiene	0.495	0.572	-15.6	105	0.00
95 T	Naphthalene	2.447	2.596	-6.1	96	0.00
96 T	1,2,3-Trichlorobenzene	0.928	1.043	-12.4	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6