

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
 Data File : VY020752.D
 Acq On : 30 Dec 2024 13:22
 Operator : SY/MD
 Sample : VY1230SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1230SBL01

Quant Time: Dec 31 00:37:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	133756	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.610	114	216108	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.414	117	185225	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.346	152	76871	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	63493	57.954	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.900%
35) Dibromofluoromethane	7.628	113	66165	55.584	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.160%
50) Toluene-d8	10.103	98	252784	53.285	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.560%
62) 4-Bromofluorobenzene	12.402	95	79267	50.108	ug/l	-0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	100.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
Data File : VY020752.D
Acq On : 30 Dec 2024 13:22
Operator : SY/MD
Sample : VY1230SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1230SBL01

Quant Time: Dec 31 00:37:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

