

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
 Data File : VY020755.D
 Acq On : 30 Dec 2024 15:17
 Operator : SY/MD
 Sample : P5391-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-12272024

Quant Time: Dec 31 00:39:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	1152	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	1793	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.414	117	875	50.000	ug/l	#-0.01
72) 1,4-Dichlorobenzene-d4	13.359	152	155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	733	77.682	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.360%
35) Dibromofluoromethane	7.634	113	737	74.625	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	149.240%#
50) Toluene-d8	10.109	98	1649	42.264	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.520%
62) 4-Bromofluorobenzene	12.408	95	113	8.610	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	17.220%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
Data File : VY020755.D
Acq On : 30 Dec 2024 15:17
Operator : SY/MD
Sample : P5391-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-12272024

Quant Time: Dec 31 00:39:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

