

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
 Data File : VY020756.D
 Acq On : 30 Dec 2024 15:40
 Operator : SY/MD
 Sample : P5390-01RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-12272024RE

Quant Time: Dec 31 00:40:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	21708	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.616	114	29330	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	19361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	5336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	7430	41.787	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.580%
35) Dibromofluoromethane	7.634	113	8416	52.094	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.180%
50) Toluene-d8	10.103	98	31692	49.354	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.408	95	6619	30.829	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	61.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
Data File : VY020756.D
Acq On : 30 Dec 2024 15:40
Operator : SY/MD
Sample : P5390-01RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-12272024RE

Quant Time: Dec 31 00:40:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

