



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123024\
Data File : VY020750.D
Acq On : 30 Dec 2024 08:48
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Title : SW846 8260
Last Update : Thu Dec 26 16:12:28 2024



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	12371	PASS
75	95	30	60	53.4	38413	PASS
95	95	100	100	100.0	71936	PASS
96	95	5	9	6.6	4756	PASS
173	174	0.00	2	0.6	346	PASS
174	95	50	100	84.6	60848	PASS
175	174	5	9	7.8	4757	PASS
176	174	95	101	99.7	60640	PASS
177	176	5	9	6.6	4010	PASS