

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123124\
 Data File : VY020760.D
 Acq On : 31 Dec 2024 10:42
 Operator : SY/MD
 Sample : VY1231SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1231SBL01

Quant Time: Jan 01 00:19:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	143183	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.616	114	234812	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	197633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	78826	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	66748	56.913	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.820%
35) Dibromofluoromethane	7.634	113	71335	55.154	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.300%
50) Toluene-d8	10.109	98	274686	53.289	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.580%
62) 4-Bromofluorobenzene	12.408	95	82544	48.023	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	96.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY123124\
Data File : VY020760.D
Acq On : 31 Dec 2024 10:42
Operator : SY/MD
Sample : VY1231SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1231SBL01

Quant Time: Jan 01 00:19:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

