

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011320\  
 Data File : VY001168.D  
 Acq On : 13 Jan 2020 09:25  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jan 14 10:30:03 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jan 10 16:01:58 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0   | 95    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.384 | 0.384 | 0.0   | 97    | 0.00     |
| 3 P   | Chloromethane               | 0.553 | 0.560 | -1.3  | 103   | 0.00     |
| 4 C   | Vinyl Chloride              | 0.558 | 0.586 | -5.0# | 100   | 0.00     |
| 5 T   | Bromomethane                | 0.300 | 0.344 | -14.7 | 106   | 0.00     |
| 6 T   | Chloroethane                | 0.322 | 0.347 | -7.8  | 103   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.776 | 0.782 | -0.8  | 100   | 0.00     |
| 8 T   | Diethyl Ether               | 0.302 | 0.302 | 0.0   | 97    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.500 | 0.516 | -3.2  | 100   | 0.00     |
| 10 T  | Methyl Iodide               | 0.684 | 0.712 | -4.1  | 100   | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.053 | 0.049 | 7.5   | 99    | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.511 | 0.529 | -3.5# | 101   | 0.00     |
| 13 T  | Acrolein                    | 0.048 | 0.046 | 4.2   | 95    | 0.00     |
| 14 T  | Allyl chloride              | 0.863 | 0.904 | -4.8  | 102   | 0.00     |
| 15 T  | Acrylonitrile               | 0.151 | 0.149 | 1.3   | 98    | 0.00     |
| 16 T  | Acetone                     | 0.123 | 0.119 | 3.3   | 99    | 0.00     |
| 17 T  | Carbon Disulfide            | 1.654 | 1.733 | -4.8  | 101   | 0.00     |
| 18 T  | Methyl Acetate              | 0.405 | 0.369 | 8.9   | 94    | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 1.414 | 1.411 | 0.2   | 97    | 0.00     |
| 20 T  | Methylene Chloride          | 0.683 | 0.579 | 15.2  | 98    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.580 | 0.605 | -4.3  | 101   | 0.00     |
| 22 T  | Diisopropyl ether           | 1.762 | 1.844 | -4.7  | 101   | 0.00     |
| 23 T  | Vinyl Acetate               | 1.127 | 1.170 | -3.8  | 99    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 0.965 | 1.009 | -4.6  | 102   | 0.00     |
| 25 T  | 2-Butanone                  | 0.191 | 0.189 | 1.0   | 97    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 0.887 | 0.919 | -3.6  | 103   | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.635 | 0.662 | -4.3  | 100   | 0.00     |
| 28 T  | Bromochloromethane          | 0.396 | 0.423 | -6.8  | 94    | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.131 | 0.128 | 2.3   | 96    | 0.00     |
| 30 C  | Chloroform                  | 0.945 | 0.985 | -4.2# | 100   | 0.00     |
| 31 T  | Cyclohexane                 | 1.021 | 1.022 | -0.1  | 100   | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 0.839 | 0.869 | -3.6  | 100   | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.528 | 0.530 | -0.4  | 93    | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 94    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.293 | 0.290 | 1.0   | 95    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.456 | 0.476 | -4.4  | 100   | 0.00     |
| 37 T  | Ethyl Acetate               | 0.267 | 0.251 | 6.0   | 92    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.418 | 0.439 | -5.0  | 100   | 0.00     |
| 39 T  | Methylcyclohexane           | 0.618 | 0.637 | -3.1  | 99    | 0.00     |
| 40 TM | Benzene                     | 1.330 | 1.394 | -4.8  | 100   | 0.00     |
| 41 T  | Methacrylonitrile           | 0.126 | 0.132 | -4.8  | 100   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.372 | 0.389 | -4.6  | 100   | 0.00     |
| 43 T  | Isopropyl Acetate           | 0.511 | 0.513 | -0.4  | 97    | 0.00     |
| 44 TM | Trichloroethene             | 0.347 | 0.368 | -6.1  | 100   | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.329 | 0.344 | -4.6# | 101   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011320\  
 Data File : VY001168.D  
 Acq On : 13 Jan 2020 09:25  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jan 14 10:30:03 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jan 10 16:01:58 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 46 T  | Dibromomethane              | 0.179 | 0.185 | -3.4  | 98    | 0.00     |
| 47 T  | Bromodichloromethane        | 0.437 | 0.454 | -3.9  | 100   | 0.00     |
| 48 T  | Methyl methacrylate         | 0.237 | 0.234 | 1.3   | 95    | 0.00     |
| 49 T  | 1,4-Dioxane                 | 0.002 | 0.002 | 0.0   | 88    | 0.00     |
| 50 S  | Toluene-d8                  | 1.128 | 1.199 | -6.3  | 95    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.259 | 0.253 | 2.3   | 94    | 0.00     |
| 52 CM | Toluene                     | 0.848 | 0.894 | -5.4# | 100   | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.489 | 0.512 | -4.7  | 99    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.552 | 0.579 | -4.9  | 100   | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.266 | 0.271 | -1.9  | 98    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.407 | 0.406 | 0.2   | 95    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.457 | 0.469 | -2.6  | 97    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.143 | 0.133 | 7.0   | 98    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.182 | 0.185 | -1.6  | 97    | 0.00     |
| 60 T  | Dibromochloromethane        | 0.306 | 0.319 | -4.2  | 98    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.260 | 0.261 | -0.4  | 96    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.461 | 0.457 | 0.9   | 94    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 93    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.322 | 0.334 | -3.7  | 97    | 0.00     |
| 65 PM | Chlorobenzene               | 0.999 | 1.049 | -5.0  | 99    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.344 | 0.369 | -7.3  | 100   | 0.00     |
| 67 C  | Ethyl Benzene               | 1.828 | 1.954 | -6.9# | 100   | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.689 | 0.751 | -9.0  | 100   | 0.00     |
| 69 T  | o-Xylene                    | 0.661 | 0.715 | -8.2  | 101   | 0.00     |
| 70 T  | Styrene                     | 1.160 | 1.248 | -7.6  | 100   | 0.00     |
| 71 P  | Bromoform                   | 0.212 | 0.222 | -4.7  | 96    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 93    | 0.00     |
| 73 T  | Isopropylbenzene            | 3.759 | 4.031 | -7.2  | 101   | 0.00     |
| 74 T  | N-amyl acetate              | 1.173 | 1.175 | -0.2  | 93    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 0.791 | 0.797 | -0.8  | 96    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 0.535 | 0.562 | -5.0  | 105   | 0.00     |
| 77 T  | Bromobenzene                | 0.831 | 0.869 | -4.6  | 98    | 0.00     |
| 78 T  | n-propylbenzene             | 4.521 | 4.843 | -7.1  | 100   | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.597 | 2.768 | -6.6  | 100   | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.136 | 3.363 | -7.2  | 99    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.316 | 0.321 | -1.6  | 96    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 2.772 | 2.981 | -7.5  | 100   | 0.00     |
| 83 T  | tert-Butylbenzene           | 2.635 | 2.892 | -9.8  | 102   | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 3.115 | 3.340 | -7.2  | 100   | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.791 | 4.028 | -6.3  | 99    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.292 | 3.531 | -7.3  | 99    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.571 | 1.661 | -5.7  | 99    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.617 | 1.704 | -5.4  | 100   | 0.00     |
| 89 T  | n-Butylbenzene              | 3.315 | 3.491 | -5.3  | 97    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011320\  
Data File : VY001168.D  
Acq On : 13 Jan 2020 09:25  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.00G/5ML/MSVOA Y/SOIL  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
LabSampleId :  
VSTDCCC050

Quant Time: Jan 14 10:30:03 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
Quant Title : SW846 8260  
QLast Update : Fri Jan 10 16:01:58 2020  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 90 T | Hexachloroethane            | 0.643 | 0.676 | -5.1 | 97    | 0.00     |
| 91 T | 1,2-Dichlorobenzene         | 1.460 | 1.540 | -5.5 | 99    | 0.00     |
| 92 T | 1,2-Dibromo-3-Chloropropane | 0.143 | 0.136 | 4.9  | 90    | 0.00     |
| 93 T | 1,2,4-Trichlorobenzene      | 0.962 | 0.941 | 2.2  | 92    | 0.00     |
| 94 T | Hexachlorobutadiene         | 0.485 | 0.476 | 1.9  | 91    | 0.00     |
| 95 T | Naphthalene                 | 2.239 | 2.129 | 4.9  | 89    | 0.00     |
| 96 T | 1,2,3-Trichlorobenzene      | 0.848 | 0.823 | 2.9  | 92    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6