

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
 Data File : VY003915.D
 Acq On : 14 Jan 2021 12:04
 Operator : SY/MD
 Sample : VY0114SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0114SBL01

Quant Time: Jan 15 00:56:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 12 11:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	148569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	219714	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	205310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	107331	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	109324	47.37	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.74%
35) Dibromofluoromethane	7.722	113	87990	49.00	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.00%
50) Toluene-d8	10.179	98	346649	47.64	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.28%
62) 4-Bromofluorobenzene	12.477	95	120379	45.30	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
Data File : VY003915.D
Acq On : 14 Jan 2021 12:04
Operator : SY/MD
Sample : VY0114SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0114SBL01

Quant Time: Jan 15 00:56:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 12 11:47:10 2021
Response via : Initial Calibration

