

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
 Data File : VY003923.D
 Acq On : 14 Jan 2021 16:51
 Operator : SY/MD
 Sample : M1100-02
 Misc : 6.25G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOPSOIL-RBF-ASHCROFT

Quant Time: Jan 15 00:58:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 12 11:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	187640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	286968	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	251779	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	119442	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	98415	33.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	67.54%
35) Dibromofluoromethane	7.722	113	83386	35.56	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	71.12%
50) Toluene-d8	10.179	98	294414	30.98	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	61.96%
62) 4-Bromofluorobenzene	12.477	95	109727	31.61	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	63.22%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
Data File : VY003923.D
Acq On : 14 Jan 2021 16:51
Operator : SY/MD
Sample : M1100-02
Misc : 6.25G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOPSOIL-RBF-ASHCROFT

Quant Time: Jan 15 00:58:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 12 11:47:10 2021
Response via : Initial Calibration

