

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
 Data File : VY003926.D
 Acq On : 14 Jan 2021 18:01
 Operator : SY/MD
 Sample : M1124-05RE
 Misc : 4.72G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-223RE

Quant Time: Jan 15 00:59:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 12 11:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	65158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	92303	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	64282	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	22228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	29811	29.45	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	58.90%
35) Dibromofluoromethane	7.716	113	25954	34.41	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	68.82%
50) Toluene-d8	10.179	98	106421	34.81	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	69.62%
62) 4-Bromofluorobenzene	12.477	95	26457	23.70	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	47.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011421\
Data File : VY003926.D
Acq On : 14 Jan 2021 18:01
Operator : SY/MD
Sample : M1124-05RE
Misc : 4.72G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-223RE

Quant Time: Jan 15 00:59:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011221S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 12 11:47:10 2021
Response via : Initial Calibration

