

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011723\
 Data File : VY012225.D
 Acq On : 17 Jan 2023 16:03
 Operator : KP/MD
 Sample : 01159-04
 Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-6

Quant Time: Jan 18 00:33:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	227530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	399741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	385050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	180716	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	129775	70.416	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.840%
35) Dibromofluoromethane	7.716	113	125853	59.459	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.920%
50) Toluene-d8	10.179	98	473655	65.969	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	131.940%
62) 4-Bromofluorobenzene	12.483	95	166716	55.874	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011723\
Data File : VY012225.D
Acq On : 17 Jan 2023 16:03
Operator : KP/MD
Sample : 01159-04
Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-6

Quant Time: Jan 18 00:33:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

