

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011723\
 Data File : VY012230.D
 Acq On : 17 Jan 2023 18:00
 Operator : KP/MD
 Sample : 01175-10
 Misc : 5.54g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 YARD-2

Quant Time: Jan 18 00:34:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	245312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	417062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	398835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	185933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	138361	69.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.180%
35) Dibromofluoromethane	7.716	113	128473	58.176	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.360%
50) Toluene-d8	10.179	98	495271	66.126	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.260%
62) 4-Bromofluorobenzene	12.483	95	169834	54.555	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011723\
Data File : VY012230.D
Acq On : 17 Jan 2023 18:00
Operator : KP/MD
Sample : 01175-10
Misc : 5.54g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
YARD-2

Quant Time: Jan 18 00:34:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

