

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011823\
 Data File : VY012249.D
 Acq On : 18 Jan 2023 16:26
 Operator : KP/MD
 Sample : 01184-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 APNB-BOT-1RE

Quant Time: Jan 19 00:36:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	14645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	24469	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	25362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	10962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	10504	89.649	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 179.300%#	
35) Dibromofluoromethane	7.716	113	7733	59.685	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 119.380%	
50) Toluene-d8	10.179	98	30003	68.432	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 136.860%	
62) 4-Bromofluorobenzene	12.483	95	10172	55.693	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.380%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	1786	2.636	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011823\
Data File : VY012249.D
Acq On : 18 Jan 2023 16:26
Operator : KP/MD
Sample : 01184-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
APNB-BOT-1RE

Quant Time: Jan 19 00:36:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

