

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012723\
 Data File : VY012383.D
 Acq On : 27 Jan 2023 15:06
 Operator : KP/MD
 Sample : 01328-15
 Misc : 7.97g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-BORING-1-14-23-GRAB-8

Quant Time: Jan 30 02:42:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	97052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	144320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	117755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	48930	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	40031	49.742	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.480%
35) Dibromofluoromethane	7.715	113	38114	49.876	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.760%
50) Toluene-d8	10.178	98	165092	63.524	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	127.040%
62) 4-Bromofluorobenzene	12.483	95	42158	39.135	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012723\
Data File : VY012383.D
Acq On : 27 Jan 2023 15:06
Operator : KP/MD
Sample : 01328-15
Misc : 7.97g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-BORING-1-14-23-GRAB-8

Quant Time: Jan 30 02:42:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

