

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013122\
 Data File : VY007342.D
 Acq On : 31 Jan 2022 16:52
 Operator : SY/MD
 Sample : N1333-07
 Misc : 3.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-7

Quant Time: Feb 01 01:53:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	114325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	213074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	204655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	88980	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	75181	53.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.180%
35) Dibromofluoromethane	7.722	113	71192	49.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.820%
50) Toluene-d8	10.185	98	261009	49.579	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.483	95	93778	50.077	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.160%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013122\
Data File : VY007342.D
Acq On : 31 Jan 2022 16:52
Operator : SY/MD
Sample : N1333-07
Misc : 3.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-7

Quant Time: Feb 01 01:53:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52




Instrument :
MSVOA_Y
ClientSampleId :
SB-7


Tgt Ion:168 Resp: 114325
Ion Ratio Lower Upper
168 100
99 59.9 49.1 73.7



#33
1,2-Dichloroethane-d4
Concen: 53.594 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. 0.006 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52

Tgt Ion: 65 Resp: 75181
Ion Ratio Lower Upper
65 100
67 52.7 0.0 105.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52




Instrument :
MSVOA_Y
ClientSampleId :
SB-7


Tgt Ion:114 Resp: 213074
Ion Ratio Lower Upper
114 100
63 21.4 0.0 39.2
88 15.7 0.0 31.0



#35
Dibromofluoromethane
Concen: 49.412 ug/l
RT: 7.722 min Scan# 968
Delta R.T. 0.006 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52

Tgt Ion:113 Resp: 71192
Ion Ratio Lower Upper
113 100
111 102.0 82.2 123.4
192 17.5 14.2 21.4









#50
Toluene-d8
Concen: 49.579 ug/l
RT: 10.185 min Scan# 11
Delta R.T. 0.007 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52




Instrument :
MSVOA_Y
ClientSampleId :
SB-7


Tgt Ion: 98 Resp: 261009
Ion Ratio Lower Upper
98 100
100 64.7 52.2 78.2



#62
4-Bromofluorobenzene
Concen: 50.077 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY007342.D
Acq: 31 Jan 2022 16:52

Tgt Ion: 95 Resp: 93778
Ion Ratio Lower Upper
95 100
174 77.3 0.0 148.4
176 74.5 0.0 146.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY007342.D

Acq: 31 Jan 2022 16:52

Instrument :

MSVOA_Y

ClientSampleId :

SB-7





Tgt Ion:117 Resp: 204655

Ion Ratio Lower Upper

117 100

82 56.8 45.4 68.0

119 31.2 26.4 39.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VY007342.D

Acq: 31 Jan 2022 16:52

Tgt Ion:152 Resp: 88980

Ion Ratio Lower Upper

152 100

115 62.6 31.4 94.0

150 158.0 0.0 348.4

