

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
 Data File : VY012558.D
 Acq On : 09 Feb 2023 11:49
 Operator : KP/MD
 Sample : 01462-20
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOT-VOC-GRAB-20

Quant Time: Feb 09 16:40:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	158775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	258660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	251467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	116049	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	88688	43.926	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.860%
35) Dibromofluoromethane	7.722	113	53923	28.626	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	57.260%
50) Toluene-d8	10.179	98	307328	39.197	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.400%
62) 4-Bromofluorobenzene	12.483	95	101399	37.077	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.160%
Target Compounds						
16) Acetone	3.960	43	1696940	3492.029	ug/l	98
29) Tetrahydrofuran	7.368	42	2401	8.237	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
Data File : VY012558.D
Acq On : 09 Feb 2023 11:49
Operator : KP/MD
Sample : 01462-20
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOT-VOC-GRAB-20

Quant Time: Feb 09 16:40:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

