

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
 Data File : VY012561.D
 Acq On : 09 Feb 2023 13:12
 Operator : KP/MD
 Sample : 01504-07
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TRIP-BLANK

Quant Time: Feb 10 05:57:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	147652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	248262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	243989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	117577	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	89215	47.516	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.040%
35) Dibromofluoromethane	7.716	113	74752	41.345	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.700%
50) Toluene-d8	10.179	98	295089	39.212	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.420%
62) 4-Bromofluorobenzene	12.483	95	98842	37.656	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.320%
Target Compounds						
29) Tetrahydrofuran	7.350	42	1832	6.759	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
Data File : VY012561.D
Acq On : 09 Feb 2023 13:12
Operator : KP/MD
Sample : 01504-07
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TRIP-BLANK

Quant Time: Feb 10 05:57:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

