

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
 Data File : VY012565.D
 Acq On : 09 Feb 2023 14:45
 Operator : KP/MD
 Sample : 01513-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-208

Quant Time: Feb 10 05:59:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	141728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	239775	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	239307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	112579	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	83758	46.474	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.940%
35) Dibromofluoromethane	7.716	113	71896	41.173	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.340%
50) Toluene-d8	10.179	98	285829	39.326	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.660%
62) 4-Bromofluorobenzene	12.483	95	94805	37.396	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
Data File : VY012565.D
Acq On : 09 Feb 2023 14:45
Operator : KP/MD
Sample : 01513-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-208

Quant Time: Feb 10 05:59:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

