

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
 Data File : VY012560.D
 Acq On : 09 Feb 2023 12:48
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 10 05:56:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	151141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	249662	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	247575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	113594	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	88329	45.958	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.920%
35) Dibromofluoromethane	7.716	113	75996	41.798	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.600%
50) Toluene-d8	10.179	98	295988	39.111	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.220%
62) 4-Bromofluorobenzene	12.483	95	97991	37.122	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.240%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020923\
Data File : VY012560.D
Acq On : 09 Feb 2023 12:48
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 10 05:56:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY012560.D
Acq: 09 Feb 2023 12:48




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 151141
Ion Ratio Lower Upper
168 100
99 52.4 40.8 61.2



#33
1,2-Dichloroethane-d4
Concen: 45.958 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. 0.000 min
Lab File: VY012560.D
Acq: 09 Feb 2023 12:48

Tgt Ion: 65 Resp: 88329
Ion Ratio Lower Upper
65 100
67 49.1 0.0 98.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012560.D

Acq: 09 Feb 2023 12:48

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 249662

Ion Ratio Lower Upper

114 100

63 17.2 0.0 33.6

88 15.1 0.0 29.8



#35

Dibromofluoromethane

Concen: 41.798 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012560.D

Acq: 09 Feb 2023 12:48




Tgt Ion:113 Resp: 75996


Ion Ratio Lower Upper

113 100

111 101.8 81.9 122.9

192 21.9 17.5 26.3









#50
Toluene-d8
Concen: 39.111 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012560.D
Acq: 09 Feb 2023 12:48




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 295988
Ion Ratio Lower Upper
98 100
100 64.7 52.2 78.2



#62
4-Bromofluorobenzene
Concen: 37.122 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012560.D
Acq: 09 Feb 2023 12:48

Tgt Ion: 95 Resp: 97991
Ion Ratio Lower Upper
95 100
174 97.2 0.0 188.0
176 93.6 0.0 183.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY012560.D

Acq: 09 Feb 2023 12:48

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion:117 Resp: 247575

Ion Ratio Lower Upper

117 100

82 50.9 41.7 62.5

119 31.5 25.8 38.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VY012560.D

Acq: 09 Feb 2023 12:48

Tgt Ion:152 Resp: 113594

Ion Ratio Lower Upper

152 100

115 56.7 28.1 84.3

150 158.9 0.0 351.0

