

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021423\
 Data File : VY012594.D
 Acq On : 14 Feb 2023 17:45
 Operator : KP/MD
 Sample : 01552-03
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-4.0-6.0-DUP

Quant Time: Feb 15 01:32:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	136060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	227022	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	227146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	106756	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	81433	47.066	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.140%
35) Dibromofluoromethane	7.716	113	67620	40.900	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.800%
50) Toluene-d8	10.178	98	270756	39.345	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.700%
62) 4-Bromofluorobenzene	12.483	95	90110	37.541	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021423\
Data File : VY012594.D
Acq On : 14 Feb 2023 17:45
Operator : KP/MD
Sample : 01552-03
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-4.0-6.0-DUP

Quant Time: Feb 15 01:32:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

