

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
 Data File : VY012603.D
 Acq On : 15 Feb 2023 10:36
 Operator : KP/MD
 Sample : VY0215SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0215SBL01

Quant Time: Feb 16 00:49:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	149077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	250278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	252079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122991	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	88603	46.739	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.480%
35) Dibromofluoromethane	7.715	113	75636	41.497	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.000%
50) Toluene-d8	10.178	98	296067	39.026	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.060%
62) 4-Bromofluorobenzene	12.483	95	101414	38.324	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012603.D
Acq On : 15 Feb 2023 10:36
Operator : KP/MD
Sample : VY0215SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0215SBL01

Quant Time: Feb 16 00:49:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

