

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
 Data File : VY012606.D
 Acq On : 15 Feb 2023 11:46
 Operator : KP/MD
 Sample : 01552-15
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-4.0-6.0

Quant Time: Feb 16 00:52:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	121913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	201862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	202066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100170	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	73796	47.602	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.200%
35) Dibromofluoromethane	7.716	113	61355	41.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.480%
50) Toluene-d8	10.179	98	238960	39.053	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	78.100%
62) 4-Bromofluorobenzene	12.483	95	80489	37.712	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012606.D
Acq On : 15 Feb 2023 11:46
Operator : KP/MD
Sample : 01552-15
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-4.0-6.0

Quant Time: Feb 16 00:52:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

