

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
 Data File : VY012612.D
 Acq On : 15 Feb 2023 14:06
 Operator : KP/MD
 Sample : 01552-09
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-05-4.0-6.0

Quant Time: Feb 16 00:54:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	177100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	168993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	77556	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	57528	40.903	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.800%
35) Dibromofluoromethane	7.710	113	50985	39.531	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.060%
50) Toluene-d8	10.179	98	208040	38.753	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.500%
62) 4-Bromofluorobenzene	12.477	95	64164	34.267	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	68.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012612.D
Acq On : 15 Feb 2023 14:06
Operator : KP/MD
Sample : 01552-09
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-05-4.0-6.0

Quant Time: Feb 16 00:54:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

