

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
 Data File : VY012617.D
 Acq On : 15 Feb 2023 16:25
 Operator : KP/MD
 Sample : 01572-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-277-281

Quant Time: Feb 16 00:58:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	137607	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	229131	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	217113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77987	44.568	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	89.140%		
35) Dibromofluoromethane	7.716	113	68235	40.892	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	81.780%		
50) Toluene-d8	10.179	98	268312	38.631	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	77.260%		
62) 4-Bromofluorobenzene	12.483	95	85050	35.107	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	70.220%		
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	19316	8.803	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012617.D
Acq On : 15 Feb 2023 16:25
Operator : KP/MD
Sample : 01572-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-277-281

Quant Time: Feb 16 00:58:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration

