

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
 Data File : VY012613.D
 Acq On : 15 Feb 2023 14:30
 Operator : KP/MD
 Sample : 01552-05RE
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-0-2.0RE

Quant Time: Feb 16 00:56:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:09:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	19793	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	29436	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	26182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	10894	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	9388	37.300	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.600%
35) Dibromofluoromethane	7.716	113	8261	38.536	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	77.080%
50) Toluene-d8	10.179	98	33221	37.232	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	74.460%
62) 4-Bromofluorobenzene	12.477	95	9310	29.914	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	59.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012613.D
Acq On : 15 Feb 2023 14:30
Operator : KP/MD
Sample : 01552-05RE
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-0-2.0RE

Quant Time: Feb 16 00:56:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:09:03 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.012 min
Lab File: VY012613.D
Acq: 15 Feb 2023 14:30




Instrument :
MSVOA_Y
ClientSampleId :
SB-03-0-2.0RE


Tgt Ion:168 Resp: 19793
Ion Ratio Lower Upper
168 100
99 53.4 40.8 61.2



#33
1,2-Dichloroethane-d4
Concen: 37.300 ug/l
RT: 8.137 min Scan# 1036
Delta R.T. -0.006 min
Lab File: VY012613.D
Acq: 15 Feb 2023 14:30

Tgt Ion: 65 Resp: 9388
Ion Ratio Lower Upper
65 100
67 48.2 0.0 98.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.006 min

Lab File: VY012613.D

Acq: 15 Feb 2023 14:30

Instrument :

MSVOA_Y

ClientSampleId :

SB-03-0-2.0RE





Tgt Ion:114 Resp: 29436

Ion Ratio Lower Upper

114 100

63 16.4 0.0 33.6

88 14.0 0.0 29.8



#35

Dibromofluoromethane

Concen: 38.536 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012613.D

Acq: 15 Feb 2023 14:30




Tgt Ion:113 Resp: 8261


Ion Ratio Lower Upper

113 100

111 101.2 81.9 122.9

192 22.2 17.5 26.3









#50
Toluene-d8
Concen: 37.232 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012613.D
Acq: 15 Feb 2023 14:30




Instrument :
MSVOA_Y
ClientSampleId :
SB-03-0-2.0RE


Tgt Ion: 98 Resp: 33221
Ion Ratio Lower Upper
98 100
100 65.1 52.2 78.2



#62
4-Bromofluorobenzene
Concen: 29.914 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY012613.D
Acq: 15 Feb 2023 14:30

Tgt Ion: 95 Resp: 9310
Ion Ratio Lower Upper
95 100
174 99.3 0.0 188.0
176 94.3 0.0 183.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY012613.D

Acq: 15 Feb 2023 14:30

Instrument :

MSVOA_Y

ClientSampleId :

SB-03-0-2.0RE





Tgt Ion:117 Resp: 26182

Ion Ratio Lower Upper

117 100

82 48.4 41.7 62.5

119 33.1 25.8 38.8



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.006 min

Lab File: VY012613.D

Acq: 15 Feb 2023 14:30

Tgt Ion:152 Resp: 10894

Ion Ratio Lower Upper

152 100

115 52.0 28.1 84.3

150 160.5 0.0 351.0

