



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021523\
Data File : VY012601.D
Acq On : 15 Feb 2023 08:53
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020723S.M
Title : SW846 8260
Last Update : Fri Feb 17 18:08:19 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	10362	PASS
75	95	30	60	51.1	33016	PASS
95	95	100	100	100.0	64568	PASS
96	95	5	9	6.5	4184	PASS
173	174	0.00	2	1.2	781	PASS
174	95	50	100	98.1	63360	PASS
175	174	5	9	7.3	4637	PASS
176	174	95	101	97.2	61557	PASS
177	176	5	9	6.9	4237	PASS