

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022123\
 Data File : VY012680.D
 Acq On : 21 Feb 2023 17:33
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 22 01:28:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	105646	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	185661	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	178870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	82903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	69079	55.364	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.720%
35) Dibromofluoromethane	7.722	113	61256	48.848	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.700%
50) Toluene-d8	10.178	98	225079	47.909	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.820%
62) 4-Bromofluorobenzene	12.483	95	77311	51.424	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.840%
Target Compounds						
						Qvalue
93) 1,2,4-Trichlorobenzene	15.007	180	1749	1.077	ug/l	98
95) Naphthalene	15.239	128	4013	1.223	ug/l #	91
96) 1,2,3-Trichlorobenzene	15.421	180	1919	1.335	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022123\
Data File : VY012680.D
Acq On : 21 Feb 2023 17:33
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 22 01:28:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

