

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
 Data File : VY013043.D
 Acq On : 22 Mar 2023 12:47
 Operator : KP/MD
 Sample : 02021-05
 Misc : 7.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-6D-032023

Quant Time: Mar 23 05:20:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	80610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	135970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	124024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	55619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	55075	68.813	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.620%	
35) Dibromofluoromethane	7.716	113	47983	58.475	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.940%	
50) Toluene-d8	10.179	98	162601	54.203	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.400%	
62) 4-Bromofluorobenzene	12.483	95	50789	50.161	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.320%	
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	11815	7.789	ug/l	91
52) Toluene	10.240	92	4216	1.797	ug/l	93
69) o-Xylene	12.026	106	3613	2.199	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
Data File : VY013043.D
Acq On : 22 Mar 2023 12:47
Operator : KP/MD
Sample : 02021-05
Misc : 7.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-6D-032023

Quant Time: Mar 23 05:20:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

