

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
 Data File : VY013050.D
 Acq On : 22 Mar 2023 15:31
 Operator : KP/MD
 Sample : 02028-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-032123

Quant Time: Mar 23 05:23:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	7755	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	11730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	11623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	5074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	6387	82.951	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 165.900%#	
35) Dibromofluoromethane	7.716	113	4949	69.911	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 139.820%	
50) Toluene-d8	10.179	98	14523	56.118	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.240%	
62) 4-Bromofluorobenzene	12.477	95	4439	50.819	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.640%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	887	3.057	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
Data File : VY013050.D
Acq On : 22 Mar 2023 15:31
Operator : KP/MD
Sample : 02028-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-032123

Quant Time: Mar 23 05:23:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

