

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
 Data File : VY013039.D
 Acq On : 22 Mar 2023 11:13
 Operator : KP/MD
 Sample : VY0322SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0322SBL01

Quant Time: Mar 23 05:17:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	93750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	152479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	136722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	65847	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	52425	56.322	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.640%
35) Dibromofluoromethane	7.716	113	52775	57.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.700%
50) Toluene-d8	10.179	98	154790	46.012	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.483	95	54779	48.244	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032223\
Data File : VY013039.D
Acq On : 22 Mar 2023 11:13
Operator : KP/MD
Sample : VY0322SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0322SBL01

Quant Time: Mar 23 05:17:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY013039.D
Acq: 22 Mar 2023 11:13




Instrument :
MSVOA_Y
ClientSampleId :
VY0322SBL01


Tgt Ion:168 Resp: 93750
Ion Ratio Lower Upper
168 100
99 52.6 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 56.322 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY013039.D
Acq: 22 Mar 2023 11:13

Tgt Ion: 65 Resp: 52425
Ion Ratio Lower Upper
65 100
67 53.6 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY013039.D

Acq: 22 Mar 2023 11:13

Instrument :

MSVOA_Y

ClientSampleId :

VY0322SBL01





Tgt Ion:114 Resp: 152479

Ion Ratio Lower Upper

114 100

63 19.4 0.0 37.4

88 15.2 0.0 28.6



#35

Dibromofluoromethane

Concen: 57.351 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY013039.D

Acq: 22 Mar 2023 11:13




Tgt Ion:113 Resp: 52775


Ion Ratio Lower Upper

113 100

111 101.7 82.2 123.2

192 20.6 16.8 25.2





#50

Toluene-d8

Concen: 46.012 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. -0.000 min

Lab File: VY013039.D

Acq: 22 Mar 2023 11:13

Instrument :

MSVOA_Y

ClientSampleId :





VY0322SBL01

Tgt Ion: 98 Resp: 154790

Ion Ratio Lower Upper

98 100

100 63.0 50.9 76.3



#62

4-Bromofluorobenzene

Concen: 48.244 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VY013039.D

Acq: 22 Mar 2023 11:13




Tgt Ion: 95 Resp: 54779


Ion Ratio Lower Upper

95 100

174 90.5 0.0 176.6

176 87.6 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013039.D
Acq: 22 Mar 2023 11:13

Instrument :
MSVOA_Y
ClientSampleId :
VY0322SBL01

Tgt Ion:117 Resp: 136722
Ion Ratio Lower Upper
117 100
82 55.7 45.4 68.0
119 31.6 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY013039.D
Acq: 22 Mar 2023 11:13

Tgt Ion:152 Resp: 65847
Ion Ratio Lower Upper
152 100
115 55.0 27.1 81.3
150 157.6 0.0 347.8

