

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032823\
 Data File : VY013120.D
 Acq On : 28 Mar 2023 16:34
 Operator : KP/MD
 Sample : 02111-05RE
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 APCS-WESTRE

Quant Time: Mar 29 02:30:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	1056	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	564	160.466	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 320.940%#	
35) Dibromofluoromethane	7.716	113	287	57.714	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.420%	
50) Toluene-d8	10.185	98	968	53.246	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.500%	
62) 4-Bromofluorobenzene	12.489	95	277	45.143	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 90.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032823\
Data File : VY013120.D
Acq On : 28 Mar 2023 16:34
Operator : KP/MD
Sample : 02111-05RE
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
APCS-WESTRE

Quant Time: Mar 29 02:30:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

