

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032823\
 Data File : VY013124.D
 Acq On : 28 Mar 2023 18:07
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 02:31:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.493	0.450	8.7	74	0.00
3 P	Chloromethane	0.589	0.506	14.1	72	0.00
4 C	Vinyl Chloride	0.540	0.520	3.7#	76	0.00
5 T	Bromomethane	0.328	0.338	-3.0	82	0.00
6 T	Chloroethane	0.324	0.333	-2.8	80	0.00
7 T	Trichlorofluoromethane	0.897	0.848	5.5	75	0.00
8 T	Diethyl Ether	0.314	0.284	9.6	69	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.525	2.8	77	0.00
10 T	Methyl Iodide	0.703	0.614	12.7	64	0.00
11 T	Tert butyl alcohol	0.070	0.048	31.4#	66	0.00
12 CM	1,1-Dichloroethene	0.515	0.492	4.5#	74	0.00
13 T	Acrolein	0.087	0.055	36.8#	50#	0.00
14 T	Allyl chloride	0.951	0.915	3.8	74	0.00
15 T	Acrylonitrile	0.166	0.171	-3.0	79	0.00
16 T	Acetone	0.195	0.171	12.3	63	0.00
17 T	Carbon Disulfide	1.753	1.496	14.7	67	0.00
18 T	Methyl Acetate	0.580	0.568	2.1	75	0.00
19 T	Methyl tert-butyl Ether	1.330	1.380	-3.8	77	0.00
20 T	Methylene Chloride	0.714	0.628	12.0	79	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.571	-2.0	79	0.00
22 T	Diisopropyl ether	1.727	1.954	-13.1	85	0.00
23 T	Vinyl Acetate	1.073	1.170	-9.0	80	0.00
24 P	1,1-Dichloroethane	0.984	1.074	-9.1	84	0.00
25 T	2-Butanone	0.241	0.226	6.2	68	0.00
26 T	2,2-Dichloropropane	0.874	0.866	0.9	78	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.606	-1.0	76	0.00
28 T	Bromochloromethane	0.360	0.358	0.6	75	0.00
29 T	Tetrahydrofuran	0.146	0.145	0.7	73	0.00
30 C	Chloroform	0.972	1.036	-6.6#	82	0.00
31 T	Cyclohexane	0.951	0.885	6.9	74	0.00
32 T	1,1,1-Trichloroethane	0.879	0.902	-2.6	79	0.00
33 S	1,2-Dichloroethane-d4	0.496	0.505	-1.8	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.302	0.329	-8.9	81	0.00
36 T	1,1-Dichloropropene	0.506	0.528	-4.3	77	0.00
37 T	Ethyl Acetate	0.319	0.325	-1.9	77	0.00
38 T	Carbon Tetrachloride	0.522	0.568	-8.8	80	0.00
39 T	Methylcyclohexane	0.601	0.614	-2.2	73	0.00
40 TM	Benzene	1.430	1.551	-8.5	79	0.00
41 T	Methacrylonitrile	0.172	0.188	-9.3	79	0.00
42 TM	1,2-Dichloroethane	0.430	0.472	-9.8	80	0.00
43 T	Isopropyl Acetate	0.569	0.581	-2.1	73	0.00
44 TM	Trichloroethene	0.377	0.389	-3.2	75	0.00
45 C	1,2-Dichloropropane	0.361	0.391	-8.3#	81	0.00
46 T	Dibromomethane	0.206	0.216	-4.9	76	0.00
47 T	Bromodichloromethane	0.485	0.534	-10.1	81	0.00
48 T	Methyl methacrylate	0.262	0.285	-8.8	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032823\
 Data File : VY013124.D
 Acq On : 28 Mar 2023 18:07
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 02:31:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	77	0.00
50 S	Toluene-d8	1.103	1.035	6.2	70	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.330	-7.1	75	0.00
52 CM	Toluene	0.863	0.939	-8.8#	79	0.00
53 T	t-1,3-Dichloropropene	0.509	0.545	-7.1	77	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.617	-7.1	77	0.00
55 T	1,1,2-Trichloroethane	0.287	0.296	-3.1	76	0.00
56 T	Ethyl methacrylate	0.382	0.402	-5.2	72	0.00
57 T	1,3-Dichloropropane	0.489	0.527	-7.8	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.216	-7.5	74	0.00
59 T	2-Hexanone	0.224	0.233	-4.0	71	0.00
60 T	Dibromochloromethane	0.349	0.375	-7.4	79	0.00
61 T	1,2-Dibromoethane	0.276	0.286	-3.6	75	0.00
62 S	4-Bromofluorobenzene	0.372	0.402	-8.1	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.364	0.374	-2.7	78	0.00
65 PM	Chlorobenzene	1.049	1.085	-3.4	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.410	-5.7	80	0.00
67 C	Ethyl Benzene	1.845	1.953	-5.9#	78	0.00
68 T	m/p-Xylenes	0.710	0.744	-4.8	77	0.00
69 T	o-Xylene	0.662	0.706	-6.6	79	0.00
70 T	Styrene	1.105	1.193	-8.0	78	0.00
71 P	Bromoform	0.253	0.271	-7.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.544	3.313	6.5	79	0.00
74 T	N-amyl acetate	1.115	0.980	12.1	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.797	0.706	11.4	77	0.00
76 T	1,2,3-Trichloropropane	0.597	0.522	12.6	75	0.00
77 T	Bromobenzene	0.861	0.782	9.2	78	0.00
78 T	n-propylbenzene	4.369	4.100	6.2	79	0.00
79 T	2-Chlorotoluene	2.428	2.280	6.1	80	0.00
80 T	1,3,5-Trimethylbenzene	2.924	2.741	6.3	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.245	14.9	72	0.00
82 T	4-Chlorotoluene	2.513	2.352	6.4	80	0.00
83 T	tert-Butylbenzene	2.524	2.451	2.9	81	0.00
84 T	1,2,4-Trimethylbenzene	2.860	2.756	3.6	81	0.00
85 T	sec-Butylbenzene	3.851	3.901	-1.3	86	0.00
86 T	p-Isopropyltoluene	3.147	3.199	-1.7	85	0.00
87 T	1,3-Dichlorobenzene	1.660	1.685	-1.5	87	0.00
88 T	1,4-Dichlorobenzene	1.694	1.656	2.2	85	0.00
89 T	n-Butylbenzene	3.013	3.019	-0.2	84	0.00
90 T	Hexachloroethane	0.655	0.641	2.1	86	0.00
91 T	1,2-Dichlorobenzene	1.501	1.520	-1.3	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.119	7.8	77	0.00
93 T	1,2,4-Trichlorobenzene	0.924	0.866	6.3	79	0.00
94 T	Hexachlorobutadiene	0.564	0.533	5.5	84	0.00
95 T	Naphthalene	1.852	1.701	8.2	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032823\
Data File : VY013124.D
Acq On : 28 Mar 2023 18:07
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 29 02:31:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.761	7.4	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6